

UREA UV SYSTEM PACK

(UV GLDH METHOD)

B Auto 200, Unicorn 230, Unicorn 120 & Bonavera Chem 200 ,
Beaconnic chem 200, Beaconnic B200, Beaconnic analyzer 120,
Bonavera chem 100(Fully Auto Biochemistry Analyzer)

Code	Product Name	Pack Size
BA233	Urea UV System Pack	4x40 + 4x10 ml

INTENDED USE

Diagnostic reagent for quantitative *in vitro* determination of Urea in human serum, plasma and urine.

CLINICAL SIGNIFICANCE

Urea is the major end product of protein nitrogen metabolism in humans. It constitutes the largest fraction of the non-protein nitrogen component of the blood. Urea is produced in the liver and excreted through the kidneys in the urine. Consequently, the circulating levels of urea depend upon protein intake, protein catabolism and kidney function. Elevated urea levels can occur with dietary changes, diseases which impair kidney function, liver diseases, congestive heart failure, diabetes and infections.

PRINCIPLE

The enzyme methodology employed in this reagent is based on the reaction first described by Talke and Schubert. To shorten and simplify the assay, the calculations are based on the discovery of Tiffany et al. that urea concentration is proportional to absorbance change over a fixed time interval.



1. Urea is hydrolysed in the presence of water and Urease to produce ammonia and carbon dioxide.
2. In the presence of Glutamate Dehydrogenase (GLDH) and reduced Nicotinamide Adenine Dinucleotide (NADH), ammonia combines with α -ketoglutarate (α -KG) to produce L Glutamate.
3. The reaction is monitored by measuring the rate of decrease in absorbance at 340 nm as NADH is converted to NAD.

REAGENT COMPOSITION

Reagent 1 : Urea Enzyme Reagent

Tris Buffer	>100 mmol/L
ADP	>1 mmol/L
Urease	>20000 U/L
GLDH	>1500 U/L
2-Oxalagutarate	>15 mmol/L

Reagent 2 : Urea Substrate Reagent

NADH	>1.05 mmol/L
------	--------------

Also contains Non-reactive fillers and stabilizers.



BEACON

STABILITY AND STORAGE

The unopened reagents are stable till the expiry date stated on the bottle and kit label when stored at +2 - +8°C.

Reagents are ready to use.

After opening, reagents are stable until expiry date at +2 - +8°C if stored at appropriate conditions, closed carefully and without any contamination.

On board stability: Min. 30 days if refrigerated (+8 - +14°) and not contaminated.

SPECIMEN COLLECTION AND HANDLING

Use serum, EDTA plasma and heparin (no ammonium heparin) plasma, urine. It is recommended to follow NCCLS procedures (or similar standardized conditions). Dilute urine 1 + 100 with dist. water and multiply results by 101.

Stability

in serum/plasma:	7 days	at +20 - +25°C
	7 days	at +4 - +8°C
	1 year	at -20°C
in urine:	2 days	at +20 - +25°C
	2 days	at +4 - +8°C
	1 month	at -20°C

Discard contaminated specimens.

CALIBRATION

Calibration with the Beacon Multicalibrator is recommended.

QUALITY CONTROL

It's recommended to run normal and abnormal control sera to validate reagent performance.

UNIT CONVERSION

mg/dl x 0.1665 = mmol/l
Urea (mg/dl) x 0.467 = BUN (mg/dl)
BUN (mg/dl) x 2.14 = Urea (mg/dl)

EXPECTED VALUES

In Serum / Plasma

10 - 40 mg/dl

Urea in Urine

26 - 43 g/24 h (0.43 - 0.72 mol/24 h)

It is recommended that each laboratory verify this range or derives reference interval for the population it serves.

PERFORMANCE DATA

Data contained within this section is representative of performance on Beacon System.

BEACON DIAGNOSTICS PVT. LTD. 424, NEW GIDC, KABILPORE, NAVSARI - 396 424. INDIA

Data obtained in your laboratory may differ from these values.

Limit of quantification: 1 mg/dl
Linearity: 300 mg/dl
Measuring range: 1-300 mg/dl

PRECISION

Intra-assay precision Within run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	42.89	1.06	2.48
Sample 2	111.96	1.62	1.45

Inter-assay precision Run to run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	24	1.12	4.60

COMPARISON

A comparison between Urea UV System Pack (y) and commercially available test (x) using 20 samples gave following results:

$$y = 1.055x + 2.825 \text{ mg/dl}$$

$$r = 0.998$$

INTERFERENCES

Following substances do not interfere:

Haemoglobin up to 7.5 g/l, bilirubin up to 30 mg/dl, triglycerides up to 2000 mg/dl.

WARNING AND PRECAUTIONS

For *in vitro* diagnostic use. To be handles by entitled and professionally educated person.

Reagents of the kit are not classified like dangerous but contains less than 0.1% sodium azide - classified as very toxic and dangerous substance for the environment. MSDS will be provided on request.

WASTE MANAGEMENT

Please refer to local legal requirements.

Parameter For B Auto 200, Unicorn 230, Unicorn 120 & Bonavera Chem200 , Beaconic chem 200, Beaconic B200, Beaconic analyzer 120, Bonavera chem 100 (Fully Auto Biochemistry Analyzer)

Test Name	UREA UV
Full Name	UREA UV
PRI Wave	340 nm
SEC Wave	630 nm
Assay/Point	FIXED TIME
Start	18
End	21
Decimal	2
Unit	mg/dl
Linearity Range Low	1
Linearity Range High	300
Sample Volume	2 µl
Reagent 1 (R1) Volume	160 µl

Reagent 1 (R2) Volume	40 µl
Substrate Depleted/Abs.limit	-
Linearity	300 mg/dl
Out Of Linearity Range	-
Calibration Type	2 Point linear
Points	2
Blank Type	Reagent
Concentration Blank	0.00
Concentration Std	Refer calibrator value sheet

NOTE









The program is made as per the in house testing, it can be modified as per requirements.

Clinical diagnosis should not be made on findings of a single test results, but both clinical and laboratory data.

REFERENCES

1. Thomas L. Clinical Laboratory Diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 374-7.
2. Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 1838.
3. Talka, H. Schubert, G.E. Kiin. Wochschr. 19; 43: 174.
4. Tiffany, T.O. Jansen, J.M Butts CA, Overton JB, Scott CD. Clin Chem 1972; 18: 829-40.
5. Kaplan, LA. in "Clinical Theory, Analysis and Correlation." Kaplan LA, Pesce AJ. (Ed) CV Mosby Company St Louis 1984; 1257-61.
6. Shephard, MD, Mezzachi, RD. Clin. Biochem. Revs. 1983; 4: 61-7.
7. Young, D.S. Effects of Drugs on Clinical Laboratory Tests. Third Edition. 1990; 21:5.
8. Wachtel, M. et al, Creation and Verification of Reference Intervals Laboratory Medicine.

Symbols Used On Labels

	Catalogue Number		Manufacturer
	See Instruction for Use		Lot Number
	Content		Storage Temperature
	Expiry Date		In Vitro Diagnostics

BEA/24/UUV/SB/IFU Ver-02
09/05/2024

